Design And Development of a Smart Patient Transfer Bed

Dr. Vinoj K[1], Sanjay Sajeevan[2], Athul Anto Jolly[3], Aswin K. Ajan[4], Aswin C. Kannan[5]
[1]*Associate Professor, Department of Mechanical Engineering, Viswajyothi College of Engineering and Technology, Vazhakulam, Ernakulam, Kerala, India*
[2]*Assistant Professor, Department of Mechanical Engineering, Viswajyothi College of Engineering and Technology, Vazhakulam, Ernakulam, Kerala, India*
[3,4,5]*Students, Department of Mechanical Engineering, Viswajyothi College of Engineering and Technology, Vazhakulam, Ernakulam, Kerala, India*

***Abstract*—This work presents the design and development of a Smart Patient Transfer Bed, an innovative solution designed to address the major challenges associated with manual patient handling in healthcare. Manual patient transfer poses a high risk of musculoskeletal injuries to caregivers and can cause significant patient discomfort or secondary injuries. The primary goal of this project is to develop a cost-effective, ergonomic, and mobile system that mechanizes the patient transfer process, thereby reducing manual labor and increasing safety. The proposed design features a sturdy mechanical system integrated with a conveyor-based transfer mechanism, an electrically powered rack-and-pinion drive for a smooth sliding mechanism, and linear actuators for precise height adjustment. System functionality is achieved through the coordinated movement of electrical components managed via an intuitive user interface, facilitating smooth, reliable, and accurate patient transfers between surfaces. In contrast to current commercial models that tend to be prohibitively expensive and underutilized in resource-constrained facilities, this project delivers an affordable, highly practical design suitable for various healthcare settings. Implementation of this prototype is anticipated to significantly decrease the physical burden on healthcare workers, enhance the comfort and dignity of patients during transfer, and offer an affordable solution to the rising demand for automation within the medical sector.**

***Index Terms*—Smart Patient Bed, Patient Transfer, Conveyor Mechanism, Rack and Pinion Drive, Cost-Effective Solution, Healthcare Technology**

I. INTRODUCTION

Healthcare systems globally seek innovative approaches to ensure the delivery of safe, efficient, and high-quality care for both patients and healthcare professionals. Among the physically demanding tasks in healthcare environments, patient transfer remains one of the most challenging. This task involves moving patients between surfaces, such as from a bed to a stretcher, wheelchair, or examination table. In many healthcare settings particularly those with limited resources manual methods remain the default approach for transferring patients who are unable to move independently, including critically ill, elderly, or mobility-impaired individuals. The complexity of patient transfer escalates significantly in emergency situations where time constraints demand rapid action without compromising safety. In these scenarios, caregivers must act swiftly and with high precision, adding further strain to an already demanding role. Manual handling under these conditions challenges the physical capabilities of healthcare workers and increases the likelihood of operational errors.

A critical concern of manual patient transfer is the high risk of injury to healthcare workers. Repeated lifting, repositioning, and transferring of patients often in awkward postures places substantial stress on the musculoskeletal system, particularly the lower back. Studies indicate that nurses and caregivers experience significantly higher rates of work-related musculoskeletal injuries compared to other professions, highlighting an urgent need for supportive technologies. This continuous physical strain leads to fatigue, reduced efficiency, and long-term health issues, ultimately impacting the overall quality of care. Patients are equally at risk during manual transfers; improper handling can result in discomfort, anxiety, pressure sores, skin breakdowns, or even fractures. This risk is heightened in elderly individuals, where minor handling disturbances can lead to severe health

complications or falls. Furthermore, manual transfer is resource-inefficient, often requiring multiple caregivers to move a single patient, which strains hospital staffing levels. To address these limitations, various mechanical and automated systems have been developed, including conveyor-based systems, adjustable beds, and powered mobility devices. Conveyor systems facilitate smooth, controlled lateral movement to eliminate sudden jerks.

Rack-and-pinion mechanisms convert rotational motion into stable linear motion, while linear actuators allow accurate height adjustments with high load capacities. The integration of mechanical components with electronic control units has yielded advanced mechatronic systems. However, most high-end automated systems are costly and require specialized maintenance, making them inaccessible to smaller healthcare facilities and developing regions. Consequently, a significant gap remains between advanced technology and its practical implementation. This project aims to bridge that gap by utilizing basic mechanical principles and practical design approaches to build a safe, cost-effective, reliable, and easily operated intermediate patient transfer device.

II. METHODOLOGY

The development process integrates systematic engineering phases encompassing design, analysis, and structural evaluation, as outlined below:

1. Problem Identification & Literature Review: Examination of current patient transfer systems to discover limitations regarding cost, ergonomics, and automation levels. Synthesis of peer-reviewed journals and patents regarding smart beds.
2. Conceptual Design: Establishing the conceptual layout encompassing both the mechanical and electrical subsystems, while defining dimensional metrics based on ergonomic needs.
3. Component Selection & Calculations: Selection of DC geared motors, linear actuators, and conveyor mechanisms alongside rigorous torque, speed, and load calculations.
4. 3D Modeling & Simulation: Development of CAD models followed by structural Finite Element Analysis (FEA) using ANSYS Workbench under uniformly distributed loads (1000 N to 1500 N).
5. Fabrication Plan & Iteration: Material selection and structural layout design for the sliding frame, motor mounts, and roller bed structure up to final verification.

Figure 1 illustrates the systematic methodology followed in the design, fabrication, and implementation of the Smart Patient Transfer Bed.




Fig. 1 Methodology of the Work

III. DESIGN AND FABRICATION




Fig. 2 3D Model of the bed

3.1 Mechanical Architecture

The core system components integrate mechanical, electrical, and structural elements into a single, compact frame.

The foundational mechanical elements include:

- Conveyor Belt: Facilitates smooth lateral patient transfer between surfaces.
- Rollers: Support the belt to maintain continuous, uniform motion and optimal weight distribution.
- Sliding Mechanism: Allows the conveyor section to extend outward to bridge the gap to an adjacent bed or stretcher.
- Slot Arrangement: Guides the sliding framework and preserves structural alignment.
- Supporting Frame: Manufactured from mild steel to maximize structural strength and stability.
- Lockable Caster Wheels: Provide secure mobility and absolute stability during transfer operations.

3.2 System Specifications

The chosen electromechanical actuators and drives are detailed below:

- Conveyor System Motor: MY1016Z DC Gear Motor (250 W, 330 RPM, 35 Nm torque, 12 V DC, 23 A rated current).
- Sliding Mechanism Motor: Low-speed, high-torque DC Gear Motor (50 W, 40 RPM, 30 Nm torque).
- Height Adjustment System: GM64 AC Motor/Actuator (220 V AC, 1400 RPM).

3.3 Design Assumptions and Calculations

Structural Load Assumptions

The maximum patient mass (m_p) is assumed to be 150 kg. The total design load (including structural margins) is 1500 N uniformly distributed over the frame. Finite Element Analysis (FEA) performed using ANSYS Workbench verified that induced stresses and deformations lie well within acceptable safety factors for mild steel, ensuring structural integrity.

Actuator Lifting Force Calculations

To verify height adjustment capabilities, a nominal operational mass configuration was evaluated: Patient weight limit = 100 kg, Bed frame weight ≈ 30 kg, yielding a Total Mass (m) = 130 kg.

The total vertical lifting force required (F) is calculated as:

$F = m \times g = 130\ kg \times 9.81\ m/s^2 = 1275.3\ N$

Since the load is shared equally between two synchronized linear actuators, the force required per actuator ($F_{required}$) is:

$F_{required} = 1275.3\ N / 2 = 637.65\ N$

The height adjustment drive features a motor spinning at 1400 RPM paired with a 32:1 gear reduction mechanism, yielding an output speed of:

$N_{out} = 1400 / 32 = 43.75 \approx 44\ RPM$

Using the standard power-torque-velocity relationship ($P = T \times \omega$), the available output torque (T) at the lead screw is approximately 11 Nm. Given a lead screw pitch radius (r) of 10 mm (0.01 m), the mechanical lifting force generated per actuator ($F_{generated}$) is computed as:

$F_{generated} \approx T / r = 11\ Nm / 0.01\ m = 1100\ N$

Safety Verification: Since $F_{generated}$ (1100 N) > $F_{required}$ (637.65 N), the selected actuator configuration provides a more than sufficient safety margin to lift the bed and patient safely.

Sliding Motor Dynamics

The sliding mechanism converts the high-torque motor rotation into a stable linear travel velocity (V = 0.04 m/s) over a sprocket/pinion radius (R = 0.03 m):

$\omega = V / R = 0.04\ m/s / 0.03\ m = 1.33\ rad/s$

Accounting for friction and movement dynamics under full load conditions (m = 120 kg):

$F = 120 \times 9.81 = 1177.2\ N$

$T = F \times r = 1177.2\ N \times 0.03\ m = 35.316\ Nm$

$P = T \times \omega = 35.316\ Nm \times 1.33\ rad/s = 46.97\ W$

The results perfectly validate the selection of the 50 W sliding motor.

Conveyor Motor and Speed Reduction Calculations

- Load Estimation: Patient weight (100 kg) + Belt/Roller assembly weight (20 kg) = 120 kg Total Load.
- Force Required: F = 120 kg × 9.81 m/s² = 1177.2 N.
- Torque Required at Roller: Given a final conveyor roller sprocket radius r = 0.03 m (28 teeth, ∅ = 6 cm)

To achieve a safe, slow patient transfer speed, a 3-stage speed reduction system is implemented:

- Stage 1 (Motor to Gearbox Input): 9-tooth motor sprocket to 18-tooth gearbox sprocket → Reduction Ratio = 0.5. Output speed N_2 = 330 × 0.5 = 165 RPM.
- Stage 2 (Internal Gearbox): Fixed reduction ratio of 1
- Stage 3 (Gearbox Output to Conveyor Roller): 12-tooth sprocket to 28-tooth roller sprocket → Reduction Ratio = 12/28 = 0.428. Final roller speed N_4 = 27.5 × 0.428 = 11.77 ≈ 12 RPM.

The total system reduction ratio achieved is:
Total Reduction = 2 × 6 × (28 / 12) = 27.96
The raw motor torque derived from the power formula ($P = 2\pi NT/60$) is:
$T_{motor} = (250\ W \times 60) / (2\pi \times 330\ RPM) = 7.23\ Nm$
Thus, the final torque available at the conveyor roller ($T_{available}$) is:
$T_{available} = T_{motor} \times \text{Total Reduction} = 7.23\ Nm \times 27.96 = 202.1\ Nm$
Performance Verification: Since $T_{available}$ (202.1 Nm) ≫ $T_{required}$ (35.31 Nm), the selected drive motor is fully capable of driving the conveyor belt under full load conditions without stalling.

3.4 Electrical Control Architecture

The electrical control system coordinates all electromechanical movements, regulating the conveyor system, sliding mechanism, and height-adjusting linear actuators. To optimize operator convenience and safety, the architecture incorporates redundant control topologies: Manual Control via tactile physical switches, and Wireless Control via integrated Bluetooth communication. The core components include an ESP32 microcontroller, a 4-channel relay module, mechanical limit switches to prevent over-extension, a Bluetooth interface, and a highly regulated power supply network.

IV. ANALYSIS

4.1. Structural Analysis Overview & Model Preparation

To assess the mechanical stability, structural integrity, and deflection characteristics of the proposed patient transfer bed during operation, static structural finite element analyses (FEA) were conducted using ANSYS Workbench 2025 R2. The three-dimensional solid geometry was imported via a standardized STEP pipeline and mapped with isotropic structural steel material properties due to its high yield strength and resilience. The core simulation objectives focused on extracting directional field parameters total displacement deformation (δ), equivalent Von-Mises stress (σ_v), and equivalent elastic strain (ε) under fixed kinematic boundary conditions. Two distinct gravitational force cases, 1000 N (nominal) and 1500 N (overload limit), were evaluated to verify structural safety factor boundaries.

4.2. Computational Meshing & Spatial Discretization

Spatial discretization of the continuum domain was executed via an automated patch-conforming mesh routine, setting a regulated global element sizing baseline of 0.1 m. This configuration generated an optimized mesh density encompassing approximately 30,135 nodes and 13,166 elements, establishing a solid balance between computational execution speed and grid independence accuracy.
Advanced curvature-based adaptive refinement parameters were enabled to increase element concentration near geometric discontinuities, while smoothing techniques minimized shape distortion to prevent localized numerical artifacts during solver matrix inversion.

4.3. Simulation Environment Setup & Boundary Conditions

Kinematic constraints were applied to represent real-world operation by defining the lower base pad and wheel mounting faces as fixed supports, nullifying all translational and rotational degrees of freedom (Δx, Δy, Δz = 0). The external forces were applied uniformly across the upper load-bearing chassis profiles, oriented vertically downward along the negative Y-axis to replicate gravitational patient loads. Forces were steadily ramped up through incremental load-stepping convergence loops inside the

Mechanical APDL solver to achieve smooth numerical equilibrium.

4.4. Numerical Stress, Strain, and Deformation Results

- Equivalent Von-Mises Stress (σ_v):

Evaluated to predict potential material yield and failure under multi-axial force states. Under the 1000 N baseline load, the maximum stress reached a minor 2.40 MPa. For the 1500 N overload configuration, the peak stress shifted linearly to 3.61 MPa. Both absolute maxima remain far below the material yield strength of structural steel ($\sigma_y \approx 250$ MPa), yielding a structural factor of safety exceeding 60.

- Equivalent Elastic Strain (ε):

Measures micro-deformation per unit length within the mechanical matrix. The analysis yielded an elastic strain of 1.23×10^{-5} m/m under the 1000 N case, scaling up to 1.85×10^{-5} m/m at 1500 N. These highly restricted strain orders confirm that the material behaves purely within its linear-elastic profile, guaranteeing zero permanent micro-yielding or residual deformation upon unloading.

- Total Deformation (δ):

Defines global displacement vectors across structural nodes to evaluate chassis stiffness. Under the 1000 N force profile, the peak recorded displacement was constrained to 2.69×10^{-4} m (0.269 mm). Under the intensive 1500 N load case, the maximum global displacement reached a negligible 4.04×10^{-4} m (0.404 mm), which is completely negligible for clinical applications and confirms high structural rigidity.

4.5. Analytical Synthesis & Structural Compliance

The numerical calculations demonstrate a predictable, stable, and linear structural response across all test states. Because the induced stress and displacement fields remain well within permissible engineering thresholds, the analysis confirms that the proposed smart transfer bed provides sufficient rigidity and load-bearing safety margins. The chassis is fully qualified to sustain both standard patient loads and transient shock spikes without the risk of physical distortion or mechanical yielding over an extended operational lifecycle.

V. RESULTS AND DISCUSSION

5.1 Experimental Observations

Table 1: Experimental Operation Parameters and Performance

Parameter	Observation	Remarks
Maximum Load Capacity	Stable and safe operation	No visible structural deformation
Conveyor Belt Speed	0.04 m/s smooth, controlled motion	Ideal for maximizing patient comfort
Chain & Sprocket Motion	Linear travel (0.04 m/s)	Minimal audible noise and vibration
Height Adjustment Range	750 - 850 mm vertical range	Highly reliable vertical alignment
Power Source Efficiency	12 V DC motor system	Low power consumption profiles

5.2 Comparative Analysis

Table 2: Engineering and Economic Comparison

Feature	Commercial Automated Beds	Proposed Designed Bed
Cost	Extremely high (5 – 10 Lakh INR)	Low-cost (Below 20,000 INR)
Transfer Mechanism	Fully Automatic	Semi-Automatic
Maintenance	Highly complex, proprietary parts	Simple, low-cost local maintenance
Safety Features	High-end electronic sensors	Dependable mechanical limit switches
Portability	Moderate	High portability and low footprint

VI. APPLICATIONS

- Hospitals, Clinics, & ICUs: Facilitates effortless transfers between beds, stretchers, and diagnostic tables without disturbing complex, connected clinical lines or patient monitoring devices.

- Emergency & Trauma Care Units: Enables swift, stabilized lateral transfers under tight clinical time constraints.
- Operating Rooms: Minimizes post-surgical body disturbances when transferring patients back to recovery beds.
- Home Healthcare & Senior Living Facilities: Minimizes the daily physical workload of single caregivers when moving bedridden or geriatric individuals.
- Diagnostic Centers: Simplifies positioning onto imaging bays, including X-ray, CT, and MRI machinery.
- Field Hospitals & Disaster Response: Offers a lightweight, highly portable, and structurally robust patient handling setup operable in off-grid locations.

VII. CONCLUSION

This project successfully designed, analyzed, and manufactured a functional prototype of a cost-effective, safe, and ergonomically optimized Smart Patient Transfer Bed. By coupling a conveyor-based lateral transfer belt, a rack-and-pinion horizontal sliding mechanism, and linear vertical lift actuators, the device eliminates the physical hazards of manual lifting.

Financially, the prototype costs under 20,000 INR, which is a tiny fraction of the cost of commercial alternatives, making automation accessible to healthcare facilities in developing countries. The structural and electrical subsystems prioritize user safety, clinical dignity, and high energy efficiency. Ultimately, this design stands as a highly scalable and practical modern engineering solution to a critical problem in the medical field.

VIII. FUTURE SCOPE

- IoT Infrastructure Integration: Implementing Internet of Things (IoT) connectivity modules will support real-time remote monitoring, systemic performance analytics, and equipment fleet tracking across large clinical facilities.
- Advanced Safety Sensing Arrays: Integrating electronic weight/load cells can automatically adapt motor speeds based on real-time patient weight metrics. Ultrasonic or infrared proximity sensors will actively prevent cross-surface collisions or structural misalignment during setup.
- Intelligent Control Algorithms: Replacing basic relays with microcontrollers using smart acceleration and deceleration algorithms can optimize power delivery and eliminate mechanical jerks across varying load conditions.
- Continuous Patient Vital Monitoring: Incorporating non-invasive biometric sensors directly within the bed framework (e.g., heart rate, core temperature, and interface pressure mapping sensors) will transform the platform into an intelligent, comprehensive patient care node.

REFERENCES

[1] L. Hrabovský, J. Blata, L. Hrabec, and J. Fries, “The detection of forces acting on conveyor rollers of a laboratory device simulating the vertical section of a Sandwich Belt Conveyor,” *Measurement*, vol. 207, Art. no. 112376, 2023.

[2] D. K. Madheswaran, R. Ganesamoorthy, R. Manikandan, R. Vezhavendhan, R. Ravi, G. Suresh, and P. Chandramohan, “Design and simulation analysis of a rack & pinion seat recliner mechanism,” *AIP Conference Proceedings*, vol. 3122, no. 1, Art. no. 100013, 2024.

[3] Z. Chen, M. Zeng, and A. Fuentes-Aznar, “Geometric design, meshing simulation and stress analysis of pure rolling rack and pinion mechanisms,” *Journal of Mechanical Design*, vol. 142, no. 3, Art. no. 031122, 2020.

[4] S. A. Enoi, R. B. Ismail, and A. B. M. Desa, “Designing a smart transfer patient bed,” in *Proc. 2015 Innovation & Commercialization of Medical Electronic Technology Conference (ICMET)*, 2015, pp. 111–113.

[5] H. Wang and F. Kasagami, “A patient transfer apparatus between bed and stretcher,” *IEEE Transactions on Systems, Man, and Cybernetics, Part B (Cybernetics)*, vol. 38, no. 1, pp. 60–67, 2008.

[6] A. Sadri, P. Datar, and S. Kar, “Design and implementation of automatic height adjustment mechanism on a general-purpose platform,” in *Proc. 2018 3rd International Conference for Convergence in Technology (I2CT)*, Pune, India, 2018, pp. 1–5.

[7] F. Song, X. Yang, W. Dong, et al., “Research and prospect of textured sliding bearing,” *The International Journal of Advanced Manufacturing Technology*, vol. 121, pp. 1–25, 2022.

[8] Z. Ye and L. Wang, “Optimization analysis on assembly interference of cylindrical roller bearings,” *Advances in Mechanical Engineering*, vol. 7, no. 7, 2015, doi: 10.1177/1687814015593868.

[9] F. I. Malek, M. Muaz, S. Rubiah, M. N. Mansor, and M. A. Muda, “Design and development to improve mechanism of conveyor system part 1,” in *Proc. 2nd Integrated Design Project Conference (IDPC)*, 2015, pp. 1–12.